




S.N.	Purchaser's Specifications	Bidder's Compliance (Yes / No)	Reference Page No.	Remarks
	X-Ray Machine 500 mA with accessories			
	Manufacturer			
	Brand			
	Type/Model			
	Country of Origin			
1	Description of Functions			
1.1	A general purpose Line Frequency X-ray machine 500mA with Manual multi position table.			
2	Operational Requirements			
2.1	It shall be suitable to be used for adult, pediatric and infant patients in general Radiography examination.			
3	System Configurations			
3.1	X-ray Generator, 1 unit.			
3.2	X-Ray tube & tube support system, 1 unit.			
3.3	Manual multi position patient table, 1 unit.			
3.4	Floor mounted Bucky stand, 1 unit.			
4	Technical Specifications			
I	X-ray Generator:			
4.1	Generator Output: not less than 40 kW			
4.2	Radiographic voltage: 40 kV to 100kV.			
4.3	Radiographic current: 50 to 500mA.			
4.4	Exposure time: approx. 0.01sec (1msec) to 5 sec in 24 steps.			
4.5	Shall come with overload protection device.			
4.6	Power supply: 3 phase, 380-415V 50/60Hz.			
II	Control Console:			
4.7	Digital display for parameters.			
4.8	Minimum 2-point exposure technique.			
4.9	Status display,			
4.10	Shall come with radiography remote control in control room.			
III	X-Ray Tube: Bidder shall indicate brand and model information here and provide technical data document for X-ray tube offered.			
4.11	The tube assembly shall be floor to ceiling mounted.			
4.12	Max. Exposure voltage: 100KV.			
4.13	Max. Exposure current: 500mA.			
4.14	Large focus not more than 2.0 mm.			
4.15	Small focus not more than 1.0 mm.			


Er. Suresh Maharjan
 Biomedical Engineer
 159 Biomedical "A"

4.14	Large focus not more than 2.0 mm.			
4.15	Small focus not more than 1.0 mm.			
4.16	Maximum tube voltage 100 KV. Maximum tube output shall match with the generator output of not less than 40 KW.			
4.17	Filtration: min. 2.5mm Al equivalent.			
4.18	Cooling method passive or forced air and/or oil cooling.			
4.19	The anode shall be a rotating anode, with a speed not less than 3000rpm.			
4.20	Anode heat capacity shall not be less than 140KHU.			
4.21	It shall come with tube overload and overheat protection.			
IV	Floor to Ceiling Tube Stand:			
4.22	Longitudinal travel: approx. 1700mm.			
4.23	Vertical travel: from in the range 600 -1900mm.			
4.24	Rotation of tube arm around vertical axis: 180°; lockable at 0° to +/- 90°.			
V	Collimator:			
4.25	Collimator with manual exposure field setting having additional filters for Dose Reduction.			
4.26	Halogen lamp.			
4.27	Horizontal Bucky Table Radiography patient table:			
4.28	Radiography table top dimension: approximately 2200L x 650 W mm with effective X-ray transparent area of approximately 1750 x 650mm.			
4.29	It shall be a radio-translucent multi position table top with load capacity of more than 200kg.			
4.30	Come with grid and cassette tray, with grid ratio: 10:1. Grid line number: approx. 40 line/cm.			
4.31	Cassette size: accept all sizes up to 35x43 cm type.			
VI	Floor Mounted Bucky Stand:			
4.32	Vertical travel: from in the range 460-1700mm.			
4.33	Moving Grid with Grid ratio not less than 10:1.			
4.34	Cassette size: accept all sizes from 5"x7" to 14"x17".			
5	Accessories, Spare Parts and Consumables			
5.1	Accessories: <input type="checkbox"/> Lead apron, light weight-01 nos.			
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).			
6	Operating Environment			
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.			


Er. Suresh Maharjan
 Biomedical Engineer
 159 Biomedical "A"

6.2	Power supply: 380-415VAC 3 phase 50Hz fitted with appropriate plug for X-ray generator.			
7	Standards & Safety Requirements			
7.1	Must submit ISO 13485:2016 and BIS, AERB Approved product or Equivalent			
7.2	CE (93/42 EEC Directives) or USFDA or BIS/AERB approved product certificate.			
8	Installation and Commissioning & User Training			
8.1	Supplier must accomplish proper installation and commissioning of the equipment onsite.			
8.2	Must provide user training (including how to use and maintain the equipment)			
9	Warranty & Maintenance Service During Warranty Period			
9.1	Comprehensive warranty for 2 years after installation			
9.2	During warranty period supplier must ensure preventive maintenance and corrective/breakdown maintenance whenever required			
10	Authorization			
10.1	Manufacturer's Authorization or Local Distributor Authorization (Manufacturer's Authorization to the main Distributor is also required in case of Local Authorization)			
11	Documentation			
11.1	User (Operating) manual in English			
11.2	Service (Technical / Maintenance) manual in English			
11.3	Certificate of calibration and inspection from factory.			
11.4	Shall accomplish all the Qualification Criteria of the bidder mentioned in the main bid document			



E. Suresh Maharjan
 Biomedical Engineer
 159 Biomedical "A"

S.N	Purchaser's Specification	Bidder's Compliance Sheet		
		Yes	No	Remarks
	5-Part Differential Hematology Analyzer			
	Manufacturer			
	Brand			
	Type/Model			
	Country of Origin			
1	Description of Function			
1.1	Automated Blood Cell Counter is used to count various types of blood cells in the blood			
2	Measured Parameters & Criteria:			
2.1	It must have 28 parameters: WBC, LYM%, MON%, NEU%, EOS%, BAS%, LYM#, MON#, NEU#, EO S#, BAS#, RBC#, HGB, HCT, MCV, MCH, MCHC, RDW_CV, RDW_SD, PLT, MPV, PD W, PCT, P_LCR, P_LCC, RETIC%, RETIC_ABS, IRF * Research Parameters : ALY#, ALY%, LIC#, LIC%, NRBC#, NRBC%			
3	System Configuration			
3.1	Automatic Cell Counter, Complete unit with all standard accessories.			
4	Technical Specification			
4.1	Measurement Principle: WBC/DIFF: Flow Cytometry, Semi-conductor laser light multi-dimensional cell classification WBC Analysis: Optical and impedance measurement RBC/PLT Analysis : Impedance method HGB Test : Cyanide-free reagent colorimetry			
4.2	Reagents : Maximum of 4 reagents Diluent, Sheath, Detergent, Lyse			
4.3	Sample volume : Whole Blood : 20µL Pre-Diluted : 20µL			
4.4	Throughput : Whole Blood : Up to 60 samples/Hr. Pre-diluted: Up to 60 samples/H			
4.5	Cylindrical disk aperture for higher resolution			
4.6	Manual and automatic dilution			
4.7	Bubble and Clog detection during counting			
4.8	RFID detection system			
4.9	Data Storage: 2,00,000 results with histograms & scattergrams.			
4.10	Should have minimum 10.4 inches Touch Screen			
5	Quality Control			
5.1	Dedicated 4 QC Menu with 9 user definable normal range			
5.2	Levey-Jennings graphs			
5.3	Aperture size WBC : 100µm, RBC/PLT : 68 µm			
6	Accessories, spares and consumables			



Er. Suresh Maharjan
Biomedical Engineer
159 Biomedical "A"

6.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specific above)			
6.2	Suitable on-line UPS with minimum of 30 minute back up with maintenance free battery			
7	Operating Environment			
7.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.			
7.2	Power supply :110-220V (AC)			
8	Standards and Safety Requirements			
8.1	Must submit ISO 9001:2015			
8.2	CE(93/42 EEC Directives) or USFDA approved product certificate			
9	User Training			
9.1	Must provide user training (including how to use and maintain the equipment)			
10	Warranty			
10.1	Comprehensive warranty for 1 year			
11	Maintenance Service During Warranty Period			
11.1	During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.			
12	Installation and Commissioning			
12.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel : any prerequisites for installation to be communicated to the purchaser in advance, in detail.			
13	Documentation			
13.1	User (Operating) Manual in English			
13.2	Service (Technical/Maintenance) manual in English			
Bidder must completely fill the Technical Specification Form (TSF). Only Yes/No/All Complies.				



Er. Suresh Maharjan
Biomedical Engineer
159 Biomedical "A"